



### ***Declaration of Conformity***

We  
3M Health Care  
hereby declare under our sole responsibility  
that the CE marked products to which this declaration relates,

3M™ Tegaderm™ + Pad Transparent Dressing with Absorbent Pad  
Or  
3M™ Tegaderm™ + Pad Film Dressing with Nonadherent Pad

Product numbers:  
3582NP, 3584NP, 3586NP, 3589NP, 3582P, 3584P, 3586P, 3589P, 3590P, 3591P

are classified,  
according to the rule of Annex IX of the Medical Device Directive 93/42/EEC,  
as Class IIa sterile devices  
and

are in accordance with Annex V of Directive 93/42/EEC  
on the approximation of the laws of the Member States concerning medical devices.

In addition, we declare that the above mentioned devices fulfil the applicable provisions of the Directive  
93/42/EEC.

This declaration is made on the basis of the quality assurance certificate CE00493 delivered by BSI, 0086

This certificate is valid for devices originating from the following sites:

3M Brookings  
601 22nd Ave. South  
Brookings, South Dakota, 57006 USA

3M Poland Site Wroclaw  
Kwidzynska 6  
51-416 Wroclaw, Poland  
EU Representative  
3M Health Care, D41453, Neuss, Germany

Signature:

*Suzanne M. Danielson*

Suzanne M. Danielson  
Regulatory Affairs and Quality Director  
Medical Division

Date:

*September 29, 2008*