

## EU Declaration of Conformity


We hereby declare under our sole responsibility that the Eccentron meets the relevant provisions of the following European Union Directives:

- Council **Directive 93/42/EEC** of 14 June 1993 concerning medical devices as amended by Directive 2007/47/EC (**MDD**)
- **Directive 2006/42/EC** of the European Parliament and of the Council of 17 May 2006 on machinery as amended by Regulation (EU) 2019/1243
- **Directive 2011/65/EU** of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (**RoHS**)

The product also complies with the applicable requirements in the Swedish law 1993:584 and the Swedish Medical Products Agency regulation LVFS 2003:11 regarding medical devices.


The Eccentron has undergone a conformity assessment procedure required by the MDD and is manufactured in harmony with the Technical Documentation compiled as defined in the relevant Directives and retained by BTE.

Product information in regard to the **Medical Device Directive 93/42/EEC**:

 <b>Manufacturer</b>	BTE Technologies 7455-L New Ridge Road Hanover, MD 21076, USA <a href="http://www.btetechnologies.com">www.btetechnologies.com</a>	Telephone: 410.850.0333 Email: <a href="mailto:Service@btetechnologies.com">Service@btetechnologies.com</a>
<b>Product Identification</b>	<b>Device Trade Name:</b> Eccentron <b>Device Name:</b> Eccentron <b>Model:</b> LE1	
<b>UDI-DI</b>	10850390007816	
<b>EMDN (CND) code</b>	Z120616 - PHYSICAL THERAPY AND REHABILITATION SYSTEMS	
<b>Intended Purpose</b>	The system is intended to be used to increase muscle strength of the lower extremities.	
<b>Device Classification (MDD)</b>	Class IIa	
<b>Classification Rule (MDD)</b>	Rule 9	
<b>Route to Compliance (MDD)</b>	Annex II of the Medical Devices Directive	
<b>Device Classification (MDR)</b>	Class IIa	
<b>Classification Rule (MDR)</b>	Rule 11	



The Technology of Human Performance

<b>CE Marking Provision</b>	CE Certificate issued by the notified body in accordance with MDD is valid until May 26, 2024. Based on the <b>Medical Device Regulation (EU) 2017/745 (MDR)</b> Article 120 §3, the Eccentron can be placed on the EU market until May 26, 2024 provided that the device <ul style="list-style-type: none"><li>• will continue to comply with the MDD,</li><li>• there will be no significant changes in the design and intended purpose, and</li><li>• the device will comply with the MDR requirements for post market surveillance, vigilance, and registration of economic operators and of devices</li></ul>	
<b>Authorized Representative<sup>1</sup></b> 	Emergo Europe Prinsessegracht 20 2514 AP, The Hague The Netherlands	Telephone: +31.70.345.8570 Emails: <a href="mailto:EmergoEurope@ul.com">EmergoEurope@ul.com</a> <a href="mailto:EmergoVigilance@ul.com">EmergoVigilance@ul.com</a>
<b>Notified Body</b>	Intertek Semko AB Torshamnsgatan 43 Box 1103 SE-164 22 Kista Sweden	Notified Body ID Number 0413 Certificate Number 41319556

The device is CE marked since 2014.

Signed for on behalf of BTE Technologies

Ewa Kaczanowska  
PRRC/Regulatory Manager  
BTE Technologies

Hanover, MD

May 20, 2021

<sup>1</sup> AR for MDD and RoHS Directives